

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.5	0.49974	0.260	± 0.3	±	4.3E-01 mm						
0 a 1	0.99973	0.270	± 0.4	±	4.3E-01 mm						
0 a 1.5	1.49968	0.320	± 0.5	±	4.8E-01 mm						
0 a 2	1.99963	0.370	± 0.6	±	4.8E-01 mm						
0 a 2.5	2.49961	0.390	± 0.7	±	4.8E-01 mm						
0 a 3	2.99952	0.480	± 0.8	±	4.8E-01 mm						
0 a 3.5	3.49943	0.570	± 0.9	±	4.8E-01 mm						
0 a 4	3.99939	0.610	± 1	±	4.8E-01 mm						
0 a 4.5	4.49931	0.690	± 1.1	±	4.8E-01 mm						
0 a 5	4.99924	0.760	± 1.2	±	4.8E-01 mm						