

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49922	0.780	± 0.8	±	4.3E-01 mm						
0 a 1	0.99925	0.750	± 1	±	4.3E-01 mm						
0 a 1.5	1.49927	0.730	± 1.2	±	4.8E-01 mm						
0 a 2	1.99929	0.710	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49931	0.690	± 1.6	±	4.8E-01 mm						
0 a 3	2.99932	0.680	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49934	0.660	± 2	±	4.8E-01 mm						
0 a 4	3.99936	0.640	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49936	0.640	± 2.4	±	4.8E-01 mm						
0 a 5	4.99937	0.630	± 2.6	±	4.8E-01 mm						