

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Profundidad / (Depth)	5 mm	5.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	10 mm	10.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	20 mm	20.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	50 mm	50.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	60 mm	60.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	80 mm	80.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	100 mm	100.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	120 mm	120.00 mm	0 µm	± 50 µm	± 2.9E+01 µm
Profundidad / (Depth)	150 mm	150.00 mm	0 µm	± 50 µm	± 2.9E+01 µm