

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.070 °C	21.9 °C	-0.170 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.059 °C	25.1 °C	0.041 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.426 °C	31.1 °C	-0.326 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.308 %HR	50.4 %HR	3.092 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	56.971 %HR	54.7 %HR	-2.271 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.155 %HR	64.7 %HR	-2.455 %HR	± 3.0 %HR	± 1.9E+00 %HR