

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.070 °C	22.0 °C	-0.070 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.059 °C	25.4 °C	0.341 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.426 °C	31.4 °C	-0.026 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.308 %HR	47.9 %HR	0.592 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	56.971 %HR	56.2 %HR	-0.771 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.155 %HR	65.0 %HR	-2.155 %HR	± 3.0 %HR	± 1.9E+00 %HR