

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.070 °C	22.5 °C	0.430 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.059 °C	25.4 °C	0.341 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.426 °C	31.7 °C	0.274 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.308 %HR	48.5 %HR	1.192 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	56.971 %HR	55.4 %HR	-1.571 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.155 %HR	65.0 %HR	-2.155 %HR	± 3.0 %HR	± 1.9E+00 %HR