

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.070 °C	22.1 °C	0.030 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.059 °C	25.0 °C	-0.059 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.426 °C	31.4 °C	-0.026 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	47.308 %HR	46.9 %HR	-0.408 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	56.971 %HR	56.5 %HR	-0.471 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.155 %HR	64.5 %HR	-2.655 %HR	± 3.0 %HR	± 1.9E+00 %HR