

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10.0042 µl	10 µl	-0.0042 µl	± 0.2 µl	±	5.1E-02 µl
24.9603 µl	25 µl	0.0397 µl	± 0.5 µl	±	5.1E-02 µl
49.9306 µl	50 µl	0.0694 µl	± 1.0 µl	±	5.2E-02 µl