

Canal No. Channel #)	Patrón (Standard)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
1	9.899 µl	10 µl	0.1009 µl	± 0.2 µl	±	5.1E-02 µl
1	25.021 µl	25 µl	-0.0210 µl	± 0.5 µl	±	5.1E-02 µl
1	49.972 µl	50 µl	0.0282 µl	± 1.0 µl	±	5.2E-02 µl
2	10.009 µl	10 µl	-0.0094 µl	± 0.2 µl	±	5.1E-02 µl
2	24.911 µl	25 µl	0.0893 µl	± 0.5 µl	±	5.1E-02 µl
2	50.082 µl	50 µl	-0.0820 µl	± 1.0 µl	±	5.1E-02 µl
3	9.880 µl	10 µl	0.1204 µl	± 0.2 µl	±	3.3E-10 µl
3	25.020 µl	25 µl	-0.0198 µl	± 0.5 µl	±	5.1E-02 µl
3	49.986 µl	50 µl	0.0143 µl	± 1.0 µl	±	5.2E-02 µl
4	10.005 µl	10 µl	-0.0050 µl	± 0.2 µl	±	5.1E-02 µl
4	24.917 µl	25 µl	0.0830 µl	± 0.5 µl	±	5.1E-02 µl
4	50.074 µl	50 µl	-0.0740 µl	± 1.0 µl	±	5.1E-02 µl
5	9.890 µl	10 µl	0.1104 µl	± 0.2 µl	±	5.1E-02 µl
5	25.019 µl	25 µl	-0.0185 µl	± 0.5 µl	±	5.1E-02 µl
5	49.987 µl	50 µl	0.0126 µl	± 1.0 µl	±	5.2E-02 µl
6	10.001 µl	10 µl	-0.0010 µl	± 0.2 µl	±	5.1E-02 µl
6	24.936 µl	25 µl	0.0640 µl	± 0.5 µl	±	5.1E-02 µl
6	50.045 µl	50 µl	-0.0450 µl	± 1.0 µl	±	5.1E-02 µl
7	9.880 µl	10 µl	0.1204 µl	± 0.2 µl	±	5.1E-02 µl
7	25.023 µl	25 µl	-0.0230 µl	± 0.5 µl	±	5.1E-02 µl
7	49.989 µl	50 µl	0.0110 µl	± 1.0 µl	±	5.2E-02 µl
8	10.003 µl	10 µl	-0.0030 µl	± 0.2 µl	±	5.1E-02 µl
8	24.915 µl	25 µl	0.0850 µl	± 0.5 µl	±	5.1E-02 µl
8	50.069 µl	50 µl	-0.0690 µl	± 1.0 µl	±	5.1E-02 µl