

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.213 °C	29.8 °C	-0.413 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.760 %HR	24 %HR	-4.760 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.323 %HR	42 %HR	-7.323 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.345 %HR	68 %HR	-7.345 %HR	± 5.0 %HR	± 1.9E+00 %HR