

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.213 °C                  | 29.9 °C            | -0.313 °C              | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 28.760 %HR                 | 31 %HR             | 2.240 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 49.323 %HR                 | 51 %HR             | 1.677 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 75.345 %HR                 | 79 %HR             | 3.655 %HR              | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |