

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.213 °C	30.0 °C	-0.213 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.760 %HR	29.9 %HR	1.140 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.323 %HR	46.9 %HR	-2.423 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.345 %HR	71.1 %HR	-4.245 %HR	± 5.0 %HR	± 1.9E+00 %HR