

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.70 μm	0.05 μm	± 2.5 μm	±	5.8E-01 μm
155.23 μm	154.0 μm	-1.23 μm	± 2.5 μm	±	5.8E-01 μm
257.12 μm	257.0 μm	-0.12 μm	± 3.8 μm	±	5.8E-01 μm
511.72 μm	511.0 μm	-0.72 μm	± 14.3 μm	±	5.8E-01 μm
952 μm	955 μm	3.0 μm	± 28.6 μm	±	6.0E-01 μm