

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.20 μm	24 μm	0.8 μm	± 5 μm	±	5.8E-01 μm
51.10 μm	52 μm	0.7 μm	± 5 μm	±	5.8E-01 μm
125.10 μm	126 μm	0.90 μm	± 6 μm	±	5.8E-01 μm
249.40 μm	250 μm	0.60 μm	± 12 μm	±	5.8E-01 μm