

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.082 °C	30.1 °C	0.018 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.292 %HR	32.4 %HR	3.108 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.036 %HR	52.1 %HR	3.064 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	71.122 %HR	73.9 %HR	2.778 %HR	± 3.0 %HR	± 1.9E+00 %HR