

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)	
49.99 mmHg	50 mmHg	0.01 mmHg	± 3 mmHg	±	6.6E-01 mmHg
99.99 mmHg	100 mmHg	0.01 mmHg	± 3 mmHg	±	6.5E-01 mmHg
149.99 mmHg	150 mmHg	0.01 mmHg	± 3 mmHg	±	6.6E-01 mmHg
199.98 mmHg	200 mmHg	0.02 mmHg	± 3 mmHg	±	6.8E-01 mmHg
249.98 mmHg	250 mmHg	0.02 mmHg	± 3 mmHg	±	7.5E-01 mmHg
299.98 mmHg	300 mmHg	0.02 mmHg	± 3 mmHg	±	7.5E-01 mmHg