

| Tipo de Medida <i>(Measurement Type)</i> | Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |
|--|--------------------------|------------------|----------------------|--------------------------|------------------------------------|
| Exterior / (Exterior) | 1.200 in | 1.200 in | 0 in | ± 0.0020 in | ± 1.0E-03 in |
| Exterior / (Exterior) | 2.400 in | 2.400 in | 0 in | ± 0.0020 in | ± 1.0E-03 in |
| Exterior / (Exterior) | 3.600 in | 3.600 in | 0 in | ± 0.0020 in | ± 1.0E-03 in |
| Exterior / (Exterior) | 4.800 in | 4.800 in | 0 in | ± 0.0028 in | ± 1.0E-03 in |
| Exterior / (Exterior) | 6.000 in | 6.000 in | 0 in | ± 0.0028 in | ± 1.0E-03 in |
| Interior / (Interior) | 0.600 in | 0.600 in | 0 in | ± 0.0020 in | ± 6.7E-04 in |
| Interior / (Interior) | 5.400 in | 5.400 in | 0 in | ± 0.0028 in | ± 6.7E-04 in |
| Profundidad / (Depth) | 0.600 in | 0.600 in | 0 in | ± 0.0020 in | ± 7.6E-04 in |
| Profundidad / (Depth) | 5.400 in | 5.400 in | 0 in | ± 0.0028 in | ± 7.6E-04 in |