

Certificado No. (Certificate #):		61520	Equipo (Instrument): INDICADOR DE DIAL		
Fecha de Calibración (Calibration Date):		2025-09-23	Marca (Brand): HUMBOLDT		
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
0.500 mm		0.500 mm	0 µm	±	1.2E+00 µm
1.000 mm		1.000 mm	0 µm	±	1.2E+00 µm
1.500 mm		1.500 mm	0 µm	±	1.2E+00 µm
2.000 mm		2.000 mm	0 µm	±	1.2E+00 µm
3.500 mm		3.500 mm	0 µm	±	1.2E+00 µm
5.000 mm		5.000 mm	0 µm	±	1.2E+00 µm
7.500 mm		7.500 mm	0 µm	±	1.2E+00 µm
10.000 mm		10.000 mm	0 µm	±	1.2E+00 µm
15.000 mm		15.000 mm	0 µm	±	1.2E+00 µm