

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.213 °C	30.0 °C	-0.213 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.323 %HR	51.9 %HR	2.577 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.435 %HR	60.2 %HR	0.765 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.233 %HR	67.7 %HR	-2.533 %HR	± 3.0 %HR	± 1.9E+00 %HR