

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
18 °C	18.230 °C	18.3 °C	0.070 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	23.053 °C	23.1 °C	0.047 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	28.157 °C	28.1 °C	-0.057 °C	± 0.5 °C	± 4.0E-01 °C
45 %HR	44.210 %HR	45.3 %HR	1.090 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.425 %HR	60.3 %HR	0.875 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.345 %HR	74.3 %HR	-1.045 %HR	± 3.0 %HR	± 1.9E+00 %HR