

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.030 °C	30.4 °C	0.370 °C	± 2.0 °C	± 9.8E-01 °C
30 %HR	29.631 %HR	31.2 %HR	1.569 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.384 %HR	48.2 %HR	-1.184 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.644 %HR	70.8 %HR	1.156 %HR	± 5.0 %HR	± 1.9E+00 %HR