

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
300.242 s	300 s	-0.242 s	-0.0806%	± 0.0011%	± 2.7E+00 s
1800.235 s	1800 s	-0.235 s	-0.0131%	± 0.0011%	± 2.7E+00 s
3600.365 s	3600 s	-0.365 s	-0.0101%	± 0.0011%	± 2.7E+00 s