

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.131 °C	16 °C	0.869 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.707 °C	22 °C	1.293 °C	± 3.0 °C	± 4.0E-01 °C
30 °C	30.213 °C	29 °C	-1.213 °C	± 3.0 °C	± 4.0E-01 °C
58 %HR	57.155 %HR	55 %HR	-2.155 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.548 %HR	62 %HR	2.452 %HR	± 3.0 %HR	± 1.9E+00 %HR
62 %HR	62.053 %HR	61 %HR	-1.053 %HR	± 3.0 %HR	± 1.9E+00 %HR

