

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.213 °C	29 °C	-1.213 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.760 %HR	33 %HR	4.240 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.323 %HR	52 %HR	2.677 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.345 %HR	75 %HR	-0.345 %HR	± 3.0 %HR	± 1.9E+00 %HR