

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 1 mm              | 1.001 mm  | 1 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 1.25 mm           | 1.251 mm  | 1 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 1.5 mm            | 1.502 mm  | 1 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 2 mm              | 2.001 mm  | 1 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 3 mm              | 3.001 mm  | 1 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 4 mm              | 4.002 mm  | 2 µm          | ± 4 µm            | ±                           | 5.9E-01 µm |
| 5 mm              | 5.001 mm  | 1 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 10 mm             | 10.001 mm | 1 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 15 mm             | 15.002 mm | 2 µm          | ± 4 µm            | ±                           | 5.8E-01 µm |
| 20 mm             | 20.001 mm | 1 µm          | ± 4 µm            | ±                           | 5.9E-01 µm |
| 25 mm             | 25.000 mm | 0 µm          | ± 4 µm            | ±                           | 5.9E-01 µm |