

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.882 °C	29.9 °C	0.018 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.607 %HR	25.0 %HR	-3.607 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	47.993 %HR	44.9 %HR	-3.093 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.151 %HR	66.7 %HR	-3.451 %HR	± 3.0 %HR	± 1.9E+00 %HR