

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.350 °C	30.0 °C	0.650 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	32.6 %HR	2.787 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	53.096 %HR	55.7 %HR	2.604 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.885 %HR	69.6 %HR	-0.285 %HR	± 3.0 %HR	± 1.9E+00 %HR