

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.551 °C	30.0 °C	0.449 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	32.6 %HR	2.787 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	53.096 %HR	55.7 %HR	2.604 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.885 %HR	69.6 %HR	-0.285 %HR	± 3.0 %HR	± 1.9E+00 %HR