

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.032 °C	31.3 °C	1.268 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	20.4 %HR	-10.170 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.284 %HR	38.3 %HR	-11.984 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	55.8 %HR	-14.443 %HR	± 5.0 %HR	± 1.9E+00 %HR