

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.100 in	0.100 in	0.000 in	± 0.001 in	±	5.8E-04 in
0.125 in	0.125 in	0.000 in	± 0.001 in	±	5.8E-04 in
0.200 in	0.200 in	0.000 in	± 0.001 in	±	5.9E-04 in
0.250 in	0.250 in	0.000 in	± 0.001 in	±	6.0E-04 in
0.300 in	0.300 in	0.000 in	± 0.001 in	±	6.1E-04 in
0.350 in	0.350 in	0.000 in	± 0.001 in	±	6.3E-04 in
0.400 in	0.400 in	0.000 in	± 0.001 in	±	6.4E-04 in
0.450 in	0.450 in	0.000 in	± 0.001 in	±	6.6E-04 in
0.500 in	0.500 in	0.000 in	± 0.001 in	±	6.8E-04 in