

Certificado No. (Certificate #):		61609		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-09-23		Marca (Brand):		Total	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		5.00 mm	5.01 mm	10 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		10.00 mm	10.01 mm	10 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		50.00 mm	50.01 mm	10 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		100.00 mm	99.99 mm	-10 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		150.00 mm	149.99 mm	-10 µm	± 30 µm	± 7.4E+00 µm	
Interior / (Interior)		25.00 mm	24.99 mm	-10 µm	± 30 µm	± 8.2E+00 µm	
Interior / (Interior)		100.00 mm	99.99 mm	-10 µm	± 30 µm	± 8.2E+00 µm	
Profundidad / (Depth)		20.00 mm	20.00 mm	0 µm	± 30 µm	± 7.1E+00 µm	
Profundidad / (Depth)		100.00 mm	100.03 mm	30 µm	± 30 µm	± 7.1E+00 µm	