

Certificado No. (Certificate #):

61610

Equipo (Instrument):

INDICADOR DE DIAL

Fecha de Calibración (Calibration Date):

2025-09-23

Marca (Brand):

Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm		±	5.8E+00 µm
2.00 mm	2.00 mm	0 µm		±	5.8E+00 µm
3.00 mm	3.00 mm	0 µm		±	5.8E+00 µm
4.00 mm	4.00 mm	0 µm		±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm		±	5.8E+00 µm
6.00 mm	6.00 mm	0 µm		±	5.8E+00 µm
8.00 mm	8.00 mm	0 µm		±	5.8E+00 µm
10.00 mm	10.00 mm	0 µm		±	5.8E+00 µm
12.00 mm	12.00 mm	0 µm		±	5.8E+00 µm