

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
25 °C	24.778 °C	25.1 °C	0.272 °C	± 0.3 °C	± 4.0E-01 °C
50 %HR	51.000 %HR	52 %HR	1.000 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	62.564 %HR	64 %HR	1.436 %HR	± 3.0 %HR	± 1.9E+00 %HR

Resultados (Results) Punto de Rocío

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.228 °C	11.92 °C	1.69 °C	± 2.5 °C	± 4.0E-01 °C
15 °C	14.850 °C	13.08 °C	-1.77 °C	± 2.5 °C	± 4.0E-01 °C