

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
25 °C	24.577 °C	25.1 °C	0.473 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	51.000 %HR	52 %HR	1.000 %HR	± 2.0 %HR	± 1.9E+00 %HR
65 %HR	62.564 %HR	64 %HR	1.436 %HR	± 2.0 %HR	± 1.9E+00 %HR

Sonda Externa

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.827 °C	10.1 °C	0.303 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.874 °C	19.9 °C	-0.004 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.919 °C	39.6 °C	-0.359 °C	± 0.5 °C	± 4.0E-01 °C