

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
8 °C	8.002 °C	8.1 °C	0.098 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	14.9 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.092 °C	30.4 °C	0.308 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	28.0 %HR	-2.468 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	50.8 %HR	0.772 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	70.1 %HR	-0.143 %HR	± 3.0 %HR	± 1.9E+00 %HR