

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
8 °C	8.002 °C	8.3 °C	0.298 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.2 °C	0.179 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.092 °C	30.1 °C	0.008 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	31.2 %HR	0.732 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	48.7 %HR	-1.328 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	71.3 %HR	1.057 %HR	± 3.0 %HR	± 1.9E+00 %HR