

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
8 °C	8.002 °C	7.8 °C	-0.202 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.0 °C	-0.021 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.092 °C	30.3 °C	0.208 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	29.3 %HR	-1.168 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	50.5 %HR	0.472 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.4 %HR	-0.843 %HR	± 3.0 %HR	± 1.9E+00 %HR