

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
8 °C	8.002 °C	7.8 °C	-0.202 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.092 °C	30.2 °C	0.108 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	28.2 %HR	-2.268 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	49.6 %HR	-0.428 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	67.8 %HR	-2.443 %HR	± 3.0 %HR	± 1.9E+00 %HR