

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
8 °C	8.002 °C	1.9 °C	-6.102 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	14.9 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.092 °C	30.3 °C	0.208 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	31.8 %HR	1.332 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	48.5 %HR	-1.528 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.3 %HR	-0.943 %HR	± 3.0 %HR	± 1.9E+00 %HR