

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 8 °C                       | 8.002 °C                   | 8.9 °C             | 0.898 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 15 °C                      | 15.021 °C                  | 14.9 °C            | -0.121 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 30.092 °C                  | 29.9 °C            | -0.192 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.468 %HR                 | 32.8 %HR           | 2.332 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.028 %HR                 | 49.8 %HR           | -0.228 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.243 %HR                 | 67.8 %HR           | -2.443 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |