

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
8 °C	8.002 °C	8.0 °C	-0.002 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.092 °C	29.8 °C	-0.292 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	28.1 %HR	-2.368 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	48.9 %HR	-1.128 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	67.9 %HR	-2.343 %HR	± 3.0 %HR	± 1.9E+00 %HR