

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	28.6 %HR	-1.766 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.233 %HR	52.3 %HR	2.067 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	71.5 %HR	1.155 %HR	± 3.0 %HR	± 1.9E+00 %HR