

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.2 °C	0.138 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	29.3 %HR	-1.066 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.233 %HR	51.3 %HR	1.067 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	71.0 %HR	0.655 %HR	± 3.0 %HR	± 1.9E+00 %HR