

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.186 °C	2.5 °C	0.314 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	4.141 °C	4.3 °C	0.159 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	6.116 °C	6.4 °C	0.284 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.473 °C	8.8 °C	0.327 °C	± 0.5 °C	± 4.0E-01 °C