

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	31.0 %HR	0.634 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.233 %HR	51.2 %HR	0.967 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	71.1 %HR	0.755 %HR	± 3.0 %HR	± 1.9E+00 %HR