

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.426 °C	20.3 °C	-0.126 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.032 °C	30.1 °C	0.068 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.2 %HR	0.630 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	71.8 %HR	1.557 %HR	± 3.0 %HR	± 1.9E+00 %HR