

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.426 °C	20.3 °C	-0.126 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.032 °C	30.4 °C	0.368 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.3 %HR	1.730 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.1 %HR	1.857 %HR	± 3.0 %HR	± 1.9E+00 %HR