

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.65 μm	0.00 μm	± 2.5 μm	±	5.8E-02 μm
155.23 μm	155.4 μm	0.17 μm	± 2.5 μm	±	6.2E-02 μm
257.12 μm	256.6 μm	-0.52 μm	± 3.8 μm	±	7.0E-02 μm
511.72 μm	512.0 μm	0.28 μm	± 14.3 μm	±	9.7E-02 μm