

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
ft	ft	in	in			ft	ft	in	in		
0 a 0.0833	0.08323	0.00080	± 0.02	±	3.2E-04 in						
0 a 0.1667	0.16645	0.00300	± 0.2	±	3.2E-04 in						
0 a 0.25	0.24968	0.00380	± 0.2	±	3.2E-04 in						
0 a 0.3333	0.33311	0.00230	± 0.2	±	3.2E-04 in						
0 a 0.4167	0.41649	0.00250	± 0.2	±	3.2E-04 in						
0 a 0.5	0.49980	0.00240	± 0.3	±	3.2E-04 in						
0 a 0.5833	0.58321	0.00110	± 0.3	±	3.2E-04 in						
0 a 0.6667	0.66659	0.00130	± 0.3	±	3.2E-04 in						
0 a 0.75	0.74993	0.00080	± 0.3	±	3.2E-04 in						
0 a 0.8333	0.83334	-0.00050	± 0.3	±	3.2E-04 in						
0 a 0.9167	0.916575	-0.00020	± 0.3	±	3.2E-04 in						
0 a 1	0.999325	0.00580	± 0.4	±	3.2E-04 in						