

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-40 °C	40.200 °C	40.1 °C	-0.100 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.212 °C	-14.9 °C	0.312 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.572 °C	-0.2 °C	0.372 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.606 °C	10.0 °C	0.394 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.5 °C	0.492 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.270 °C	40.8 °C	0.530 °C	± 0.5 °C	± 4.0E-01 °C